

Disclaimer Notice – Please Read

This component should only be installed by SPECTRO authorized service personnel. Failure to do so could result in serious injury and/or instrumentation malfunction and will void all warranties or guarantees."

Carry out only those maintenance tasks that is described in the chapter "Maintenance" of the operating instructions. The operating personnel is not permitted to proceed with any other maintenance task. Any other services (e.g., technical maintenance) or repairs must only be performed by service personnel authorized by the instrument manufacturer.

The safety measures must be checked following every technical maintenance or repair activity. The checking procedure is carried out compliant with DGUV 3 or an equivalent national regulation.

Only genuine spare parts must be used when repairing the instrument. The use of non-genuine spare parts will make the CE Declaration of Conformity null and void.